

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993
CONCERNING MEDICAL DEVICES**



MANUFACTURER: NAME: Guangdong Transtek Medical Electronics Co.,Ltd.
ADDRESS: Zone A, No.105 ,Dongli Road ,Torch Development District,
Zhongshan,Guangdong,China

MEDICAL DEVICE: BLOOD PRESSURE MONITORS: LS801A

CLASSIFICATION - ANNEX IX: CLASS IIA, RULE 10

CONFORMITY ASSESSMENT ROUTE: MDD ANNEX V

WE, THE MANUFACTURER, HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES
MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE
93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC.
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER **CE 0123**

(EC) CERTIFICATE(S): NO.G2 15 02 82800 012




EUROPEAN REPRESENTATIVE: MDSS-MEDICAL DEVICE SAFETY SERVICE GMBH
SCHIFFGRABEN, 41, 30175, HANNOVER, GERMANY

START OF CE-MARKING: 2012-12-11

PLACE, DATE OF DECLARATION: ZHONGSHAN, 2015-3-31

SIGNATURE:

NAME: 
POSITION: Vice President

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES

Standards applied:

Risk management	EN ISO 14971:2012
Labeling	EN 15223: 2013
User manual	EN 1041: 2008
General requirements for safety	EN 60601-1: 2006/IEC 60601-1:2005+A1: 2012 EN 60601-1-11:2011
Non-invasive sphygmomanometers General requirements	EN 1060-1:1995+A2:2009 EN 1060-3:1997+A2:2009 IEC/EN 80601-2-30:2009
Electromagnetic compatibility	EN 60601-1-2:2007
Usability	EN 60601-1-6:2010 & EN 62366:2007
Software life-cycle	EN 62304:2006+AC: 2008
Biological evaluation	EN ISO 10993-1:2009 EN ISO 10993-5:2009 EN ISO 10993-10:2010
Clinical Investigation	MEDDEV.2.7.1: 2009 EN 1060-4: 2004