

|  |              |  |
|--|--------------|--|
|  | <b>Title</b> | <b>EC Declaration of Conformity</b>                        |
|  | <b>File</b>  | GL control solution low medium high_EC DoC_Beurer_20220420 |

|   |   |
|---|---|
| Manufacturer:   | Beurer GmbH (see address in footer)   |
| Product category:   | Control solution for Beurer blood glucose measurement devices   |
| Product type:   | Low, medium, high   |
| The product specified above is in conformity with the following specifications.   |   |
| 98/79/EC  | In vitro medical device directive (IVDD)  |
| EDMA code and name:   | 21-06-01 Blood glucose meter  |
| Classification:   | Annex II/list B   |
| Conformity assessment procedure:  | Annex IV section 3  |
| The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate: |   |
| Certificate no. and validity:   | D1311700053, valid to 2024-12-17  |
| This declaration of conformity is issued under the sole responsibility of the manufacturer.   |   |
| Signed for and on behalf of:  | Beurer GmbH   |
| Place, date of issue:   | Ulm, 2022-04-20   |
| Name, function, signature, stamp:   | Daniel Kämmerer, Team Lead Regulatory Affairs   |
|   | <br><b>Beurer GmbH</b><br><small>Soeflinger Straße 218 • 89077 Ulm</small> |