

| | | |
|--|--------------|-------------------------------------|
|  | Title | EC Declaration of Conformity |
| | Date | 2025-02-03 |

| | |
|---|--|
| Manufacturer: | Beurer GmbH (see address in footer) |
| SRN: | DE-MF-000005422 |
| Product category: | Lancing Device |
| Product type: | LD 01 LD 02 LD 03 LD 04 |
| Intended use: | The lancing device, in combination with a separate lancet needle, is intended for taking a blood sample for measuring glucose levels in human capillary blood. Use the lancing device only on the skin areas intended for taking the glucose measurement (fingertips). |
| The product specified above is in conformity with the following specifications. | |

| | |
|----------------------------------|--|
| (EU) 2017/745 | Medical device regulation (MDR) |
| Basic-UDI-DI: | LD 01: 4211125LD01N8 LD 02: 4211125LD02NA LD 03: 4211125LD03NC LD 04: 4211125LD04NE |
| Classification/applied rule(s): | Class I/rule 1 |
| Conformity assessment procedure: | not applicable for class I devices |
| Certificate no. and validity: | D1311700056, valid to 2026-11-29 |

| | |
|------------|--|
| 2011/65/EU | Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS) EN IEC 63000:2018 |
|------------|--|

| | |
|---|---|
| This declaration of conformity is issued under the sole responsibility of the manufacturer. | |
| Signed for and on behalf of: | Beurer GmbH |
| Place, date of issue: | Ulm, 2025-02-03 |
| Name, function, signature, stamp: | Werner Meternek, Director Quality Management & Regulatory Affairs ppa.   |