

Manufacturer: Beurer GmbH (see address in footer)
SRN: DE-MF-000005422
Product category: CPR devices
Product type: RH 112
Intended use: The LifePad® motivates and supports first-aiders in cardiopulmonary resuscitation (CPR) of individuals over 12 years of age after a sudden cardiac arrest (SCA) through an acoustic and visual feedback system to evaluate the compressions performed. The LifePad® is intended for single use only and must not be used multiple times under any circumstances.

The product specified above is in conformity with the following specifications.

(EU) 2017/745 Medical device regulation (MDR)
Basic-UDI-DI: 4211125RH112EY
Classification/applied rule(s): Class I/rule 13
Conformity assessment procedure: not applicable for class I devices
Certificate no. and validity: D1311700056, valid to 2026-11-29

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)
EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH
Place, date of issue: Ulm, 2025-02-03
Name, function, signature, stamp: Werner Meternek, Director Quality Management & Regulatory Affairs

ppa. 

Beurer GmbH
Söflinger Straße 218 • 89077 Ulm